

**GOVT. OF ASSAM**  
**OFFICE OF THE SUPERINTENDENT**  
**SILCHAR MEDICAL COLLEGE HOSPITAL, SILCHAR::14, ASSAM**

NO. SMCH/ 3063

SILCHAR, DATED: 13/03/23

**NOTICE INVITING QUOTATION**

Sealed quotation affixing court fee stamp of Rs.8.25 (Rupees Eight And Twenty Five Paise) is invited on urgent basis for supply of **Wheel Chair & Patient Stretcher Trolley** for Silchar Medical College and Hospital, Silchar as and when required for a period of six months or till regular tender is floated for the same. Specifications of required items are enclosed herewith as per **Annexure A**. The quotations should be supplied in the following tabular form given below.

Sl. No.	Item	Make	Model	Quantity	Total amount in digit including all taxation and delivery charges	Total amount in words including all taxation and delivery charges
1	Wheel Chair			01		
2	Patient Stretcher Trolley			01		

Quotation must accompany the following:

1. PAN of firm/proprietor
2. Valid Trade license for mentioned items
3. GST registration
4. Income tax return for the Financial year 2021-22
5. Cancelled cheque

The supplier should be able to supply the above items within 2 days of getting supply order. Quotation envelop should be super-scribed as "**Quotation for Wheel Chair & Patient Stretcher Trolley for SMCH, Silchar**". Bid document/Annexure A can be obtained from the office of the Principal cum Chief Superintendent Silchar Medical College and Hospital, Silchar or from the official website of SMCH, Silchar <https://www.smcassam.org>.

Successful bidder should show a sample of bidded items to Tender & Purchase committee before receiving supply order.

The quotation should be dropped in the tender/quotation box of the Office of the Principal cum Chief Superintendent, SMCH, Silchar up to **11AM, 17/03/2023** and will be opened on the same day at 11:30AM.

*Signature*  
13/03/23

Superintendent  
Silchar Medical College Hospital

Memo no. SMCH/ 3064-67  
Copy to:

1. Principal cum-Chief Superintendent, SMCH, Silchar.
2. Chairman, Tender and purchase Committee, SMCH, Silchar
3. FA, SMCH, Silchar
4. Office file

Silchar, dated: 13/03/23

*Signature*  
13/03/23

Superintendent  
Silchar Medical College and Hospital

## Annexure-A

### A. Attendant Controlled Non Powered Wheel Chair with Fixed Seat Height

Back rest with Cushion: Yes  
Cushion on Seat: With  
Leg guard type: Fixed  
Slope of Seat (Degree) : 5  
Type of Back rest: Fixed  
Overall Length (mm) : 1120  
Foldable: No  
Types of Wheel Chairs: Attendant Controlled Non Powered  
Frame tube thickness (mm) : 1.2  
Clearance of frame from floor (mm) : 110  
Overall Height (mm) : 920  
Loading Capacity (Kg) : 130  
Arm rest Type: Fixed  
Cushion Material Thickness (mm) : 65  
Seat Depth (mm) : 430  
Distance between seat and foot rest (mm) : 440  
Clearance of footrest from floor (mm) : 110  
Finish of Aluminium casts: Buffing  
Hand Rim: With  
Handle for push/pull action of Wheel Chair: Yes  
Thickness of Seat (mm) : 1.2  
Slope of the backrest with respect to floor (Degree) : 95  
Rear wheel tyre type: Solid  
Arm rest height from seat (mm) : 375  
Cushion on Arm rest: With  
Types of foot rest: Swinging without Height Adjustment  
Warranty (Years) : 2 minimum  
Frame tube Dia (mm) : 25.4  
Width of wheel chair: Fixed  
Seat Height: Fixed  
Seat height from floor at the front (mm) : 500  
Thickness of back rest (mm) : 1.2  
Painting of steel components: Powder Coated

*Mud*  
11/3/2023

*[Signature]*  
11/3/23

*[Signature]*  
11/3/23

*[Signature]*  
11/3/23

*[Signature]*  
11/3/23

## B. Patient Stretcher Trolley


### Standards

Conformity to standards CE  
Conformity to Manufacturers quality Standards ISO 9001 & ISO 13485

### Performance Parameters

Overall Length of trolley in mm with 5% tolerance: 2030  
Overall width of trolley in mm with 5% tolerance: 560  
Material of Stretcher Trolley: MS  
Provision of mattress: No  
Thickness of mattress in inch: Not provided  
Length of stretcher in mm with  $\pm 5\%$  tolerance: 1830  
Width of stretcher in mm with  $\pm 5\%$  tolerance: 555  
Provision of height adjustment: No  
Standard Height: 3 feet from the floor  
Provision of Saline rod: Yes  
Material of saline rod: MS  
Finish of saline rod: Painted  
Provision of storage tray: No  
Number of castor wheels: 4  
Diameter of castor wheels in mm: 150  
Number of Lockable castor wheels: 2  
Material of Castor housing and wheels: High grade non floor-staining synthetic materials with integrated thread guards  
Wheel centre having precision ball bearing to run smoothly: Yes  
Provision of Oxygen cylinder arrangement: Yes  
Provision of railings: No  
Material of railing: Not provided  
Type of railing: Not provided  
Coating on all MS parts: Pre-treated & powder coated  
Polish on all SS parts: finished with Matt polish  
All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free: Yes  
Load carrying capacity in kgs: 140  
Provision of safety belts: No  
Material of handle: MS  
Urine Bag Holder: Yes  
Warranty: minimum 2 years

**Note:** Tender & Purchase committee has the right to reject the L1 bidder, if their items are not in conformity with specifications prescribed in Annexure A. in this situation committee has the right to offer the contract to the next successful lowest bidder.

  
11.3.23  
11/3/2023

  
11/3/23

  
11/3/23

  
11/3/23